

མནན་པའི་སྒོག་རྒྱུ་གི་གནད་ཚད།

BHUTAN STANDARD

Compressed Oxygen Gas Specifications



ICS 71.100.20

© Copyright 2021

BHUTAN STANDARDS BUREAU

The National Standards Body of Bhutan

THIMPHU 11001

མནན་པའི་ཕྱི་གཏུང་གི་གནད་ཚད

BHUTAN STANDARD

Compressed Oxygen Gas Specifications

PDF disclaimer

This PDF may contain embedded typefaces. In accordance with Adobe's Licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The Bhutan Standards Bureau accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

COPYRIGHT PROTECTED DOCUMENT

© BSB 2021

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from BSB at the address below in the country of the requester.

Director General
Bhutan Standards Bureau
Rijug Lam
Thimphu-11001
Tel: 00975-2-325104/325401
Fax: 00975-2-323712/328298
Web: www.bsb.gov.bt
Published in Thimphu, Bhutan

NATIONAL FOREWORD

This Bhutan Standard which is identical with IS 309: 2005 Compressed Oxygen Gas Specifications issued by the Indian Standards Bureau (IS) was adopted by Bhutan Standards Bureau by sub-committee on Medical and industrial gases (TC 05/SC 01) and Pharmaceutical and Traditional Medicines Technical Committee (TC 05) and approved by the Bhutan Standards Bureau Board (BSB Board) on, 2021.

The text of the IS Standard has been approved as suitable for publication as Bhutan Standard without deviation. Certain conventions are however, not identical to those used in Bhutan Standard.

Attention is particularly drawn to the following:

- a) Where the words “IS Standard” appear referring to this standard, they should be read as “Bhutan Standard”.
- b) Wherever page numbers are quoted, they are “IS (IS Standard)” page numbers.

For test methods, manufacturers may refer Annex I of this standard.

भारतीय मानक
सम्पीडित ऑक्सीजन गैस — विशिष्टि
(चौथा पुनरीक्षण)

Indian Standard

COMPRESSED OXYGEN GAS — SPECIFICATION
(*Fourth Revision*)

ICS 71.100.20

© BIS 2005

BUREAU OF INDIAN STANDARDS
MANAK BHAVAN, 9 BAHADUR SHAH ZAFAR MARG
NEW DELHI 110002

September 2005

Price Group 4

FOREWORD

This Indian Standard (Fourth Revision) was adopted by the Bureau of Indian Standards, after the draft finalized by the Industrial Gases Sectional Committee had been approved by the Chemical Division Council.

This standard was originally published in 1956 and then subsequently revised in 1965, 1974 and 1992 respectively. In third revision, a new method of test for determination of moisture has been incorporated. The requirements for packing and sampling have also been modified. The packing of compressed oxygen gas in cylinders and the marking of the cylinders has been made to conform with the provision of the *Gas Cylinder Rules*, 1981 of the Government of India, with such modifications as may be issued from time to time by the Chief Controller of Explosives.

In this standard method of test for determination of carbonmonoxide, carbondioxide and total hydrocarbons have been incorporated with a view to harmonize it with corresponding British Standard BS 4364 : 1993 'Specification for industrial oxygen'.

At present, there is no ISO Standard on the subject.

The composition of the Committee responsible for the formulation of this standard is given at Annex F.

For the purpose of deciding whether a particular requirement of this standard is complied with, the final value, observed or calculated, expressing the result of a test or analysis, shall be rounded off in accordance with IS 2 : 1960 'Rules for rounding off numerical values (*revised*)'. The number of significant places retained in the rounded off value should be the same as that of the specified value in this standard.

Indian Standard

COMPRESSED OXYGEN GAS — SPECIFICATION (Fourth Revision)

1 SCOPE

1.1 This standard prescribes requirements and methods of sampling and test for compressed oxygen gas for industrial use.

1.2 This standard does not cover aviation oxygen and that for medical or inhalation purposes.

2 REFERENCES

The standards listed below contain provisions, which through reference in this text constitute provisions of this standard. At the time of publication, the editions indicated were valid. All standards are subject to revision and parties to agreements based on this standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below:

IS No.	Title
1070 : 1992	Reagent grade water (<i>third revision</i>)
1260 (Part 1) : 1973	Pictorial marking for handling and labelling of goods: Part 1 Dangerous goods (<i>first revision</i>)
4379 : 1981	Identification of contents of industrial gas cylinders (<i>first revision</i>)
7062 : 1973	Glossary of terms used in gas industry

3 TERMINOLOGY

For the purpose of this standard, the definitions given in IS 7062 shall apply.

4 REQUIREMENTS

4.1 Purity

The purity of compressed oxygen gas shall be minimum 99.0 percent by volume when tested by the method prescribed in Annex A.

4.2 Dew Point

The dew point of oxygen gas shall be at least -40°C measured at atmospheric pressure when tested by the method prescribed in B-2, B-3 and B-4 or moisture

content in oxygen shall be not more than 0.102 g/Nm^3 corresponding to dew point of -40°C when tested by the method prescribed in B-5.

4.3 Volume Fraction of Carbonmonoxide and Carbondioxide

Volume fraction of carbonmonoxide shall not be more than 2×10^{-6} and carbondioxide shall not be more than 3×10^{-6} when determined by the method prescribed in Annex C.

4.4 Volume Fraction of Hydrocarbons

Volume fraction of hydrocarbons expressed as methane shall not be more than 50×10^{-6} when determined by the method presented in Annex D.

NOTE — Oxygen gas produced by air separation process shall not normally contain any carbonmonoxide, carbondioxide and hydrocarbon. So for presence of these impurities in industrial oxygen produced by air separation process, tests shall be carried out only on agreement of such tests between the buyer and the seller. For oxygen gas produced by other methods, determination of these impurities, that is, carbonmonoxide, carbondioxide and total hydrocarbon shall be carried out along with other tests mentioned in 4.1 and 4.2.

5 PACKING AND MARKING

5.1 Packing

The gas shall be supplied compressed in cylinders of approved designs and of suitable capacity, and conforming to the requirements prescribed in the *Gas Cylinder Rules*, 1981 of the Government of India, with such modifications as may be ordered from time to time by the Chief Controller of Explosives, Government of India, or any other duly constituted authority.

5.1.1 The quantity of oxygen gas packed in a gas cylinder shall be measured at 15°C and 760 mm of Hg and shall be expressed in cubic metre.

5.2 Marking

The marking, painting, labelling and transport of cylinders shall be in accordance with the requirements of the *Gas Cylinder Rules*, 1981 of the Government of India, with such modification as may be ordered from time to time by the Chief Controller of Explosives, Government of

India, or any other duly constituted authority. The cylinders shall also be marked as shown in IS 1260 (Part 1) and IS 4379.

5.3 BIS Certification Marking

The cylinders may also be marked with the Standard Mark.

5.3.1 The use of the Standard Mark is governed by the provisions of the *Bureau of Indian Standards Act, 1986* and Rules and Regulations made thereunder. The details

of conditions under which the licence for use of the Standard Mark may be granted to manufacturers or producers may be obtained from the Bureau of Indian Standards.

6 SAMPLING

The method of drawing representative samples of the material and criteria for determining conformity of the material to the requirements of this standard shall be as prescribed in Annex E.

ANNEX A (Clause 4.1)

DETERMINATION OF PURITY OF OXYGEN GAS

A-1 QUALITY OF REAGENTS

Unless specified otherwise, pure chemicals and distilled water (see IS 1070) shall be employed in tests.

NOTE — 'Pure chemicals' shall mean chemicals that do not contain impurities, which affect the results of analysis.

A-2 APPARATUS

The apparatus shall be as shown in Fig. 1.

A-3 REAGENTS

A-3.1 Copper Wire

A-3.1.1 Ammonia Solution — Dissolve 550 g of ammonium chloride (NH_4Cl) in 1085 ml of water and to this solution add 917 ml of ammonium hydroxide solution (relative density 0.90).

A-4 PROCEDURE

A-4.1 Invert the copper wire container *C*, remove the bung and fill it with copper wire in spiral or other form. Then fill it with ammonia solution and replace the bung. Return *C* to upright position. Pour the ammonia solution into the levelling tube *A* until it is about three quarters full. Then operate the three-way stopcocks *E* and *H* so that burette *B* is in communication with the atmosphere through inlet *E*, and by raising levelling tube *A*, completely fill burette *B*. Keeping *A* raised, turn *E*, so that *B* is in communication with *C*, and allow the solution to pass into *C*, until *D* is half full. Close *E*, lower *A* and operate *E* to draw from *C* into *B*. Close *E* when capillary of *E* is full of liquid. Raise *A* and operate *E* to expel gas from *B* to the atmosphere through *F*, closing *E* when the

capillary *H* is full of liquid. Fill the lute *G* with a head of about 50 mm of water and connect *F* with sample gas and purge the gas from the lute *G*. Draw in the sample through *F* (taking care that no air is sucked in through *F* during this operation), until the liquid level in *B* is at the zero mark and is at the same level with that in *A*. Pass the gas into *C* by suitably operating the cock *B*, and raising *A*. Shake thoroughly and then draw back the unabsorbed gas into *B* by lowering *A* and operating the cock *E*, until a little solution from the capillary trickles down into *B* in order to ensure that the capillary in the plug of cock *E* is filled with liquid before leveling up and reading the volume of the gas. Repeat the process a number of times until no further absorption takes place.

A-4.2 Renewal of the Solution

When the solution is spent, close *E*, remove the bung from *C* and drain out the solution from *C* and *D*. Invert *C* and fill it with fresh solution. Replace the bung and set *C* and *D* in proper places and in upright position. Now add fresh solution to the partially spent solution of *A* and repeat the process as indicated under A-4.1.


A-5 CALCULATION

$$\text{Oxygen, percent by volume} = \frac{V - V_1}{V} \times 100$$

where

V = volume of the gas sample taken for the test, in ml;
and

*V*₁ = volume of the gas after absorption, in ml.



- A = LEVELLING TUBE
- B = GRADUATED BURETTE FOR OXYGEN TESTING
- C = COPPER WIRE CONTAINER
- D = RESERVOIR FOR SOLUTION
- E&H = THREE-WAY STOP COCKS
- F = INLET FOR GAS SAMPLE
- G = LUTE

FIG. 1 ASSEMBLY OF APPARATUS FOR DETERMINATION OF OXYGEN

ANNEX B

(Clause 4.2)

DETERMINATION OF WATER VAPOUR

B-1 GENERAL

Presence of moisture in oxygen gas shall be determination by dew point or by weight of moisture content. For dew point determination, electrolytic hygrometer, frost or dew point hygrometer, and capacitance hygrometer shall be used. For determination of weight of moisture content, absorption method shall be followed. The procedure to be followed for determining the dew point by hygrometer method shall depend upon type of apparatus to be used and manufacturer's instruction.

B-2 ELECTROLYTIC HYGROMETER

The method is based on the absorption and electrolysis of the water vapour present in the sample gas. The electrolytic current given a direct measurement of water vapour present in the gas flowing through the instrument at a steady rate. The exact procedure to be followed shall depend on the type on apparatus to be used. The instrument manufacturer's instructions in this regard shall be followed.

B-3 FROST OR DEW POINT HYGROMETER

A metal surface on the hygrometer is cooled so that dew or frost is formed from the water vapour content of the

gas at a particular pressure which may be observed optically in the apparatus. The temperature at which the dew or frost is formed is a measure of water vapour content of the gas. The exact procedure to be followed shall depend upon the type of apparatus to be used. The instrument manufacturer's instructions in this regard shall be followed.

B-4 CAPACITANCE HYGROMETER

The method is based on the change of capacitance of the sensor when a sample gas containing water vapour passes through it. The change in capacitance gives a direct measurement of water vapour present in the gas. The procedure to be followed shall depend upon the type of apparatus to be used. The instrument manufacturer's instructions in this regard shall be followed.

B-5 ABSORPTION METHOD

B-5.1 Apparatus

The apparatus consists of the following parts assembled as shown in Fig. 2.

B-5.1.1 Gas Meter, accurate to 1 percent.




FIG. 2 ASSEMBLY OF APPARATUS FOR THE DETERMINATION OF MOISTURE

B-5.1.2 Absorption Train, three U tubes containing phosphorus pentoxide connected in series to the gas meter. The one near to the gas meter will serve as a guard to prevent moisture from backing into the first two tubes.

B-5.2 Procedure

B-5.2.1 Absorption of moisture from a known volume of oxygen gas by anhydrous calcium chloride (CaCl_2) or phosphorous pentoxide (P_2O_5) or Silica Gel or mol. Sieve held in a series of U tubes.

B-5.2.2 Quantity of gas flow through U tube train shall be measured by a rotometer or any other type of flow meter having accuracy not less than ± 5 percent. Flow rate shall be about 500 l/h and not more to ensure complete absorption of oxygen in the hygroscopic chemicals held in U tube train.

B-5.2.3 U tube train filled with hygroscopic chemicals shall be initially heated to 100°C or more depending on the chemical for 1 h to drive away any moisture absorbed previously by the chemicals. The two ends of the U tube train shall be firmly closed with rubber stopper not to allow ingress of atmospheric air. Weight of the whole U tube train shall be taken on a sensitive chemical balance of accuracy not less than ± 1 percent. The U tube train shall be heated again in two intervals for 15 min each and weighed again after each 15-minutes heating to see that there is no further fall in weight of the U tube train, ensuring that moisture from the chemicals have been removed.

B-5.2.4 At least one cylinder full of oxygen gas at full filling pressure having 6 to 7 m^3 of oxygen gas shall be connected to the U tube train through pressure regulator and flow meter. The entire quantity of oxygen gas from the cylinder shall be passed through U tube train at a rate of about 500 l/h and when the cylinder is empty in about 10 h. Both ends of the U tube train shall be firmly closed with rubber stopper not to allow any moisture from atmospheric air to get in.

B-5.2.5 U tube train after absorption of moisture from oxygen gas shall be weighed in chemical balance and weight of moisture absorbed from oxygen gas shall be found by the difference in weight of the U tube train

before commencement of flow of oxygen and after completion of the flow of oxygen.

B-5.3 Calculation

B-5.3.1 Volume of total oxygen gas flow shall be checked from the flow meter. Alternatively total quantity of oxygen gas flow can be found out from water capacity of gas cylinder punched on the body of the cylinder and the difference between the initial full pressure of oxygen gas in the cylinder and the final pressure of oxygen in the gas cylinder when flow is stopped. The quantity of gas flow can be found out in the following manner.

The water capacity of gas cylinder : X litre
punched on the body

Pressure of the gas cylinder full : P_1 bar
with oxygen

Pressure in oxygen gas cylinder when : P_2 bar
gas flow from cylinder is stopped

Total quantity of oxygen gas flow from
cylinder to the 'U' tube train:

$$= \frac{X \times (P_1 - P_2)}{1\,000} \text{ m}^3$$

'm³' shall be converted to 'N m³' by
applying gas law :

$$\frac{PV}{T} = K$$

If chemical balance employed for weighment is not enough sensitive, in that case gas from more than one oxygen cylinder may be passed to increase absorption of more moisture and thereby to increase the weight of the U tube train for easier weighment.

B-5.3.2 The weight of moisture absorbed in the U tube train in grams divided by the volume of the oxygen passed, in Nm^3 , shall show moisture content, in g/Nm^3 .

B-5.3.3 The above test shall be repeated 3 times and the average moisture content in g/Nm^3 shall be determined. This shall not be more than 0.102 g/Nm^3 corresponding to -40°C dew point.

ANNEX C

(Clause 4.3)

DETERMINATION OF VOLUME FRACTION OF CARBONMONOXIDE AND CARBONDIOXIDE

C-1 The method for determination of volume fraction of carbonmonoxide and carbondioxide present in oxygen shall use the principle of infrared absorption in specific wave length by different molecules. ' Scanning infrared

spectrometer' shall be used after calibration. Instruction of the manufacturers of the instrument shall be followed for calibration and calculation of volume fraction of carbonmonoxide and carbondioxide in oxygen sample.

ANNEX D

(Clause 4.4)

DETERMINATION OF VOLUME FRACTION OF HYDROCARBONS

D-1 Gas chromatograph with flame ionization detector, gas sampling valve, suitable valve to 'backflash' hydrocarbon to detector shall be used for the determination of volume fraction of hydrocarbon. Nitrogen gas shall

be used as carrier gas. Instruction of the manufacturer of the instrument shall be followed for calibration and calculation of volume fraction of hydrocarbons in oxygen gas.

ANNEX E

(Clause 6)

SAMPLING OF COMPRESSED OXYGEN GAS

E-1 PROCEDURE

E-1.1 On regular production, two samples selected at random shall be tested per hour of compression.

E-1.1.1 The production shall be declared as conforming

to this standards, if the sample passes the oxygen and moisture content test.

E-1.1.2 However, if the sample fails in the oxygen and moisture content test, then the sampling frequency shall be doubled till all the samples pass for consecutive 2 h.

ANNEX F

(Foreword)

COMMITTEE COMPOSITION

Industrial Gases Sectional Committee, CHD 6

<i>Organization</i>	<i>Representative(s)</i>
Praxair India Pvt Ltd, Bangalore	SHRI R. S. DHULKHED (<i>Chairman</i>)
All India Industrial Gases Manufacturers' Association, New Delhi	DR P. L. BHATIA SHRI B. N. QANUNGO (<i>Alternate</i>)
Association of Automobiles Manufacturers of India, New Delhi	REPRESENTATIVE
Bharat Heavy Electricals Ltd, Hyderabad	SHRI N. V. S. RAMANI SHRI A. RAJAMOHAN (<i>Alternate</i>)
BOC India Ltd, Kolkata	SHRI P. AMARNATH
Central Electronics Ltd, Ghazibabad	SHRI M. I. ALAM
Chief Controller of Explosives, Nagpur	SHRI R. H. BHALEKAR SHRI S. K. BHARDWAJ (<i>Alternate</i>)
Department of Electronics, New Delhi	DR U. C. PANDEY
Department of Industrial Development, New Delhi	SHRI P. K. JAIN SHRI N. C. TIWARI (<i>Alternate</i>)
Everest Kanto, Mumbai	REPRESENTATIVE
Gujrat State Fertilizer Corporation Ltd, Vadodai..	SHRI M. C. SHAH SHRI R. C. MATHUR (<i>Alternate</i>)
Indian Space Research Organization, Shriharikota	REPRESENTATIVE
Industrial Oxygen Co Ltd, Mumbai	SHRI R. R. SINGANPORIA
Mahalasa Gases & Chemicals Pvt Ltd, Bangalore	SHRI A. N. PRABHU PROF M. G. SUBBA RAO (<i>Alternate</i>)
M/s Mohan Meakin Ltd, Ghaziabad	DR E. K. JAYANARAYANAN
Ministry of Defence (DGQA), Kanpur	SHRI Y. V. VAILURE SHRI SUJIT GHOSH (<i>Alternate</i>)
National Physical Laboratory, New Delhi	REPRESENTATIVE
National Test House, Kolkata	DR B. B. PAL
Ordnance Factory, Bhandara	DR S. K. DESHMUKH SHRI V. SRINIVASAN (<i>Alternate</i>)
Rashtriya Chemicals & Fertilizers Ltd, Mumbai	DR R. N. TRIVEDI
Semi Conductor Complex Ltd, Punjab	SHRI S. S. N. PRASAD
Steel Authority of India, New Delhi	SHRI A. NAGARAJAN SHRI S. G. K. MURTHY (<i>Alternate</i>)
Steel Furnace Association of India, Surat	SHRI M. K. GUPTA
Sylvania & Laxman Ltd, New Delhi	SHRI S. K. KAPOOR SHRI NAVEEN SIKKA (<i>Alternate</i>)
The Asiatic Oxygen and Acetylene Co Ltd, Kolkata	SHRI B. N. SINHA
The Industrial Gases Ltd, Kolkata	SHRI T. GARG SHRI L. R. GARG (<i>Alternate</i>)
BIS Directorate General	DR U. C. SRIVASTAVA, Director & Head (Chem) [Representing Director General (<i>Ex-officio Member</i>)]

Member Secretary
SHRI P. GHOSH
Director (Chem), BIS

Bureau of Indian Standards

BIS is a statutory institution established under the *Bureau of Indian Standards Act, 1986* to promote harmonious development of the activities of standardization, marking and quality certification of goods and attending to connected matters in the country.

Copyright

BIS has the copyright of all its publications. No part of these publications may be reproduced in any form without the prior permission in writing of BIS. This does not preclude the free use, in the course of implementing the standard, of necessary details, such as symbols and sizes, type or grade designations. Enquiries relating to copyright be addressed to the Director (Publications), BIS.

Review of Indian Standards

Amendments are issued to standards as the need arises on the basis of comments. Standards are also reviewed periodically; a standard along with amendments is reaffirmed when such review indicates that no changes are needed; if the review indicates that changes are needed, it is taken up for revision. Users of Indian Standards should ascertain that they are in possession of the latest amendments or edition by referring to the latest issue of 'BIS Catalogue' and 'Standards : Monthly Additions'.

This Indian Standard has been developed from Doc : No. CHD 6 (1163).

Amendments Issued Since Publication

Amend No.	Date of Issue	Text Affected

BUREAU OF INDIAN STANDARDS

Headquarters :

Manak Bhavan, 9 Bahadur Shah Zafar Marg, New Delhi 110 002
Telephones : 2323 01 31, 2323 33 75, 2323 94 02

Telegrams : Manaksanstha
(Common to all offices)

Regional Offices :

	Telephone
Central : Manak Bhavan, 9 Bahadur Shah Zafar Marg NEW DELHI 110 002	{ 2323 76 17 2323 38 41
Eastern : 1/14 C.I.T. Scheme VII M, V. I. P. Road, Kankurgachi KOLKATA 700 054	{ 2337 84 99, 2337 85 61 2337 86 26, 2337 91 20
Northern : SCO 335-336, Sector 34-A, CHANDIGARH 160 022	{ 260 38 43 260 92 85
Southern : C.I.T. Campus, IV Cross Road, CHENNAI 600 113	{ 2254 12 16, 2254 14 42 2254 25 19, 2254 23 15
Western : Manakalaya, E9 MIDC, Marol, Andheri (East) MUMBAI 400 093	{ 2832 92 95, 2832 78 58 2832 78 91, 2832 78 92
Branches : AHMEDABAD. BANGALORE. BHOPAL. BHUBANESHWAR. COIMBATORE. FARIDABAD. GHAZIABAD. GUWAHATI. HYDERABAD. JAIPUR. KANPUR. LUCKNOW. NAGPUR. NALAGARH. PATNA. PUNE. RAJKOT. THIRUVANANTHAPURAM. VISAKHAPATNAM.	

Annex I

Gas Detector Tubes (Alternate Test Methods)

Gas detector tubes are cylindrical, sealed tubes consisting of an inert transparent material and are constructed to allow the passage of gas. They contain reagents adsorbed onto inert substrates that are suitable for the visualization of the substance to be detected and, if necessary, they also contain preliminary layers and/or adsorbent filters to eliminate substances that interfere with the substance to be detected. The layer of indicator contains either a single reagent for the detection of a given impurity or several reagents for the detection of several substances (monolayer tube or multilayer tube).

The test is carried out by passing the required volume of the gas to be examined through the indicator tube. The length of the colored layer or the intensity of a color change on a graduated scale gives an indication of the impurities present.

The calibration of the detector tubes is verified according to the manufacturer's instructions.

Procedure: Examine according to the manufacturer's instruction or proceed as follows:

The gas supply is connected to a suitable pressure regulator and needle valve. Connect the flexible tube fitted with a Y-piece to the valve and adjust the flow of gas to be examined to purge the tubing in order to obtain an appropriate flow (Figure 3). Prepare the indicator tube and fit to the metering, following the manufacturer's instructions. Connect the open end of the indicator tube to the short leg of the tubing and operate the pump by the appropriate number of strokes to pass a suitable volume of gas to be examined through the tube. Read the value corresponding to the length of the colored layer or the intensity of the color on the graduated scale. If negative result is achieved, indicator tubes can be verified with a calibration gas containing the appropriate impurity.

In view of the wide variety of available compressor oils, it is necessary to verify the reactivity of the oil detector tubes for the oil used. Information on the reactivity for various oils is given in the leaflet; the tube manufacturer must verify the reactivity and if necessary, provide a tube specific for this oil.




Figure 3: Apparatus for gas detector tube

Arsine detector tube

Sealed glass tube containing adsorbent filters and suitable supports for the gold salt or other appropriate indicator. The minimum value indicated is 0.25 ppm or less with a relative standard deviation of at most 20 percent.

Carbon Dioxide Detector Tube

Sealed glass tube containing adsorbent filters and suitable supports for hydrazine and crystal violet indicators. The minimum value indicated is 100 ppm with a relative standard deviation of at most 15 per cent.

Carbon Monoxide Detector Tube

Sealed glass tube containing adsorbent filters and suitable supports for di-iodine pentoxide, selenium dioxide and fuming sulfuric acid indicators. The minimum value indicated is 5 ppm or less, with a relative standard deviation of at most 15 per cent.

Hydrogen sulfide detector tube

Sealed glass tube containing absorbent filters and suitable supports for an appropriate lead salt indicator. The minimum value indicated is 0.2 ppm or less, with a relative standard deviation of at most 10 per cent.

Nitrogen Monoxide and Nitrogen Dioxide Detector Tube

Sealed glass tube containing adsorbent filters and suitable supports for an oxidizing layer [Cr (VI) salt] and the diphenylbenzidine indicator. The minimum value indicated is 0.5 ppm with a relative standard deviation of at most 15 per cent.

Oil Detector Tube

Sealed glass tube containing adsorbent filters and suitable supports for the sulfuric acid indicator. The minimum value indicated is 0.1 mg/m³ with a relative standard deviation of at most 30 percent.

Phosphine detector tube

Sealed glass tube containing absorbent filters and suitable supports for the gold salt or other appropriate indicator. The minimum value indicated is 0.2 ppm or less, with a relative standard deviation of at most 20 per cent.

Sulfur Dioxide Detector Tube

Sealed glass tube containing adsorbent filters and suitable supports for the iodine and starch indicator. The minimum value indicated is 0.5 ppm with a relative standard deviation of at most 15 per cent.

Water Vapour Detector Tube

Sealed glass tube containing adsorbent filters and suitable supports for the magnesium perchlorate indicators. The minimum value indicated is 67 ppm or less, with a relative standard deviation of at most 20 per cent.

SUB COMMITTEE ON MEDICAL & INDUSTRIAL GASES (TC 05/ SC 01)

Organization

Representative(s)

Menjong Sorig Pharmaceuticals Corporation Ltd.,
Thimphu

Mr. Sherab Tenzin (Chairman)

Drug Regulatory Authority

Ms. Jambay Wangmo

M/s Aha Oxygases Pvt. Ltd, Balujhora,
Pasakha.

Mr. Udai Kumar Poudyel

M/s Quality Gases Pvt. Ltd, Industrial Estate,
Pasakha.

Mr. Bishnu Gurung

National Drug Testing Laboratory,
Dept. of Public Health, Ministry of Health

Ms. Sonam Choden

Member Secretary

Ms. Cheten Zangmo
Standardization Division
Bhutan Standards Bureau

**PHARMACEUTICAL AND TRADITIONAL MEDICINES TECHNICAL COMMITTEE
(TC 05)**

Organization

Representative(s)

Menjong Sorig Pharmaceuticals Corporation Ltd.,
Thimphu

Mr. Sherab Tenzin (Chairman)

Essential Medicines and Technology Division,
Ministry of Health

Mr. Dechen Chopel

National Biodiversity Centre, MoAF

Mr. Chencho Dorji

Office of Consumer Protection,
Department of Trade, MoEA

Mr. Tshering Wangchuk

Medical Supplies and Procurement Division,
DOMSHI, Ministry of Health

Mr. Jangchub Peljor

National Drug Testing Laboratory,
Dept. of Public Health, Ministry of Health

Mr. Kelzang Wangdi

Quality Assurance and Standardization Division,
Ministry of Health

Mr. Kuenzang Dorji

Drug Regulatory Authority

Ms. Jambay Wangmo)

Bhutan Standards Bureau

Mr. Sherab Tenzin
Director General
(Ex-officio member)

Member Secretary

Cheten Zangmo
Standardization Division
Bhutan Standards Bureau